

Contractor:

Purchase Order #:

Client: Address:

City & State: **Date Received: Date Calibrated:**

Temperature Range: Pressure Range:

Relative Humidity Range:

Air Density:

Traceable Report #: **NIST Certificate #:**

Tested By: Procedure:

Contractor Req Recall Date:

Primary Standard Calibration Date:

Description of Weights:

Although there are two NIST numbers, one or both may apply.

K Scales of South Dakota &

West River Scale

1701 W Madison St Ste 100 Sioux Falls, SD 57104 5723

PO13643

K-Scale LLC

1701 W Madison St

Ste 100

Sioux Falls, SD 57104

09 JUL 2013 15 JUL 2013 21.18 to 21.47 °C

737.9 to 738.1 mmHg

49 to 51 %

1.1576 to 1.1589 mg/cm3

2024306

681/280058-10, 822/278785-10

12

Modified Substitution (WI05-0023)

2 Years

02/11/11, 10/21/09

Due: 02/11/15, 10/21/13

1 g - 100 g Polished Kit, ASTM Class "3", S/N 0VC3



Nominal Value	ld.	Conventiona As Found (mg)	Il Mass Corr. As Left (mg)	Unc. K=2 (mg)	Tol. (mg)	Balance Used	Standard Set Used Calibrated/due MM-DD-YY/MM-DD-YY		Assumed Density (g/cm³)
1 g		0.0356	0.0356	0.0020	0.10	327Q	K594Q	05-07-13/08-07-13	7.95
2 g		0.0431	0.0431	0.0029	0.13	327Q	K594Q	05-07-13/08-07-13	7.95
3 g		0.0315	0.0315	0.0040	0.15	327Q	K594Q	05-07-13/08-07-13	7.95
5 g		-0.0295	-0.0295	0.0059	0.18	327Q	K594Q	05-07-13/08-07-13	7.95
10 g		0.063	0.063	0.012	0.25	676Q	K594Q	05-07-13/08-07-13	7.95
20 g		0.048	0.048	0.013	0.35	676Q	K594Q	05-07-13/08-07-13	7.95
30 g		0.146	0.146	0.019	0.45	1183Q	K594Q	05-07-13/08-07-13	7.95
50 g		-0.012	-0.012	0.024	0.60	1183Q	K594Q	05-07-13/08-07-13	7.95
100 g		0.086	0.086	0.043	1.0	1183Q	K594Q	05-07-13/08-07-13	7.95

Check with your local state agency for certification of compliance on legal-for-trade items.

Prepared By:





Page 1 of 1 Page

15 JUL 2013 Dated Debbie Schieffer Lab Sup

RLWS // ass Palue CERTIFICATE

Contractor:

Purchase Order #:

Client: Address:

City & State:
Date Received:
Date Calibrated:
Temperature Range:

Pressure Range: Relative Humidity Range:

Air Density:

Traceable Report #: NIST Certificate #:

Tested By: Procedure:

Contractor Req Recall Date:

Primary Standard Calibration Date:

Description of Weights:

Although there are two NIST numbers, one or both may apply.

K Scales of South Dakota &

West River Scale

1701 W Madison St Ste 100 Sioux Falls, SD 57104 5723

PO13643 K-Scale LLC

1701 W Madison St

Ste 100

Sioux Falls, SD 57104

09 JUL 2013 15 JUL 2013 21.18 °C

737.9 mmHg

51 %

1.1588 mg/cm³ 2024306A

681/280058-10, 822/278785-10

12

Modified Substitution (WI05-0023)

2 Years

02/11/11, 10/21/09

Due: 02/11/15, 10/21/13

100 g Polished Weight, ASTM Class "3", S/N 06-J17124-22



Nominal Value	ld.	Conventiona As Found (mg)	Il Mass Corr. As Left (mg)	Unc. K=2 (mg)	z Tol.	Balance Used	Standard Set Used Calibrated/due MM-DD-YY/MM-DD-YY	Assumed Density (g/cm³)
100 g	100 g		-0.569	0.043	1.0	1183Q	K594Q 05-07-13/08-07-13	7.85

Check with your local state agency for certification of compliance on legal-for-trade items.

Prepared By:



TEL: 715-234-9171 • FAX: 715-234-6967

230 West Coleman Street • Rice Lake, WI 54868 • USA



Page 1 of 1 Page

Dated 15 JUL 2013
Debbie Schieffer Lab Supervisor

CERTIFICATE

Contractor:

Client:

Address:

City & State:

Date Received:

Date Calibrated:

Pressure Range:

Temperature Range:

Traceable Report #:

NIST Certificate #:

Relative Humidity Range:

Purchase Order #:

K Scales of South Dakota &

West River Scale

1701 W Madison St Ste 100 Sioux Falls, SD 57104 5723

PO13643

K-Scale LLC

1701 W Madison St

Ste 100

Sioux Falls, SD 57104

09 JUL 2013

15 JUL 2013

21.39 °C

738.1 mmHg

50 %

1.1583 mg/cm3

2024306B

681/280058-10, 822/278785-10

12

Modified Substitution (WI05-0023)

2 Years

02/11/11, 10/21/09

Due: 02/11/15, 10/21/13

200 g Polished Weight, ASTM Class "3", S/N 07-463352

Tested By: Procedure:

Air Density:

Contractor Req Recall Date:

Primary Standard Calibration Date:

Description of Weights:

Although there are two NIST numbers,

one or both may apply.

		Conventional Mass Corr.		Unc.			Standard Set Used	Assumed
Nominal Value	ld.	As Found (mg)	As Left (mg)	K=2 (mg)	101.	Balance Used	Calibrated/due MM-DD-YY/MM-DD-YY	Density (g/cm³)
200 g		-0.095	-0.095	0.070	2.0	619Q	K594Q 05-07-13/08-07-13	7.85

Check with your local state agency for certification of compliance on legal-for-trade items.

Prepared By:



TEL: 715-234-917.1 • FAX: 715-234-6967



Page 1 of 1 Page

Dated Debbie Schieffer Lab Supe

Mass Value Certificate

RICE LAKE WEIGHING