

**STERILITY AND GROWTH TEST CERTIFICATE FOR
BACTERIOLOGICAL MEDIUM**

1. Medium identification **COVAC'S SOLUTION** Date of manufacture **29 JUN.17**

Lot no. **131** Expiration date **29 DEC.17**

Base manufacturer **SIGMA** Base lot no. **Ao239249**


2. QUALITY CONTROL RESULTS

Performance: the following microorganisms were tested and gave the expected results:

	Microbial Species	REACTION
1	E. coli ATCC No. 25922	PURPLE
2	S. typhimurium ATCC No. 14028	YELLOW

We hereby declare that the above results for this batch satisfy
our quality control specifications.

Approved by: L.Chernovsky Manager Quality Control

Data: 18.07.17 Signed: 

: