

CERTIFICATE OF ANALYSIS AND QUALITY

Product:	ChromoStreak: UriSelect3/ MacConkey	Catalogue No.:	BD- 915
-----------------	-------------------------------------	-----------------------	---------

Batch No.:	1915	UriSelect3 Batch No.:	5G0134	Exp.:	07.2018
Manufacturing Date:	01.FEB.16				
Expiry Date:	01.JUN.2016	MacConkey Batch No.:	5084716	Exp.:	01.2020

1. APPEARANCE:

Number of units examined	50	Batch size:	4750
---------------------------------	----	--------------------	------

Physical damage (bubbles, particles, leakage, discolor, dried agar)	No. of defective units
	-

2. pH MEASUREMENT:

Agar	Measured	Specification
UriSelect3	7.39	7.3±0.2
MacConkey	7.27	7.1±0.2

3. KIT PERFORMANCE*:

Inoculum (CFU/ml)	No. of colonies	
	UriSELECT 3	MacConkey
10 ³	3	2
10 ⁴	13	11
10 ⁵	>50	>50

* Number of *E.coli* colonies at different concentrations.


4. CULTURAL RESPONSE:**

Organism	UriSELECT 3		MacConkey
	Growth	Colour	
<i>E.coli</i> ATCC 25922	luxuriant	purple	Growth
<i>P.mirabilis</i> ATCC 12453	luxuriant	brown	Growth
<i>E.faecalis</i> ATCC 29212	luxuriant	bright, frank blue	Partial or complete inhibition
<i>S.aureus</i> ATCC 25923	luxuriant	white	Inhibition

** Cultural characteristics observed after an incubation of 24 h at 35 ± 2°C.

5. STERILITY TESTING:

A statistical sample of product was monitored for sterility at 35 ± 2°C for at least 5 days and was found to be compliant with the sterility specification for the product.

Performed by:	M.Botvinnik	Date:	15.02.2016	Signed:	
Approved by:	L. Chernovsky	Date:	15.02.2016	Signed:	