



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

***Altamaha Laboratories a subsidiary of
Tindall Enterprises, Inc.***

829 S.W. Central Avenue, Blackshear, GA 31516

*(Hereinafter called the Organization) and hereby declares that Organization is accredited
in accordance with the recognized International Standard:*

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the
operation of a laboratory quality management system
(as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Biological and Environmental Testing
(As detailed in the supplement)***

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President/Operations Manager

Initial Accreditation Date:

February 27, 2013

Issue Date:

March 4, 2019

Expiration Date:

March 31, 2021

Accreditation No.:

73811

Certificate No.:

L19-122

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a
continuous accreditation cycle. The validity of this certificate should be
confirmed through the PJLA website: www.pjllabs.com*



Certificate of Accreditation: Supplement

Altamaha Laboratories a subsidiary of Tindall Enterprises, Inc.

829 S.W. Central Avenue, Blackshear, GA 31516
Contact Name: Ana Rocks Phone: 912-449-0999

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT			
Biological ^F	Potable Water	Total Coliform	SM 9221 D/E	Presence/Absence			
		Fecal Coliform					
		E.coli	SM 9221 F				
	Environmental Non-Potable Water	Total Coliform Fecal Coliform E. coli	Total Coliform	SM 9223 B	D.L. = 1 MPN/100 mL		
			E.coli	Enzyme Substrate			
			Enterococci	IDEXX Enterolert/SM 9230			
	Environmental Swabs	Total Coliform E. coli	Total Coliform	SimPlate Color Indicator	D.L. = 1 CFU/mL		
			E. coli	AOAC 2005.03			
		Total Plate Count	SimPlate Color Indicator	AOAC 2002.07			
		Staphylococcus Aureus	Petrifilm Staph Express Plate	AOAC 2003.07			
		Yeast & Mold	5-Day Petrifilm Plate Count	AOAC 997.02			
		Shiga Toxin E.coli	Assurance GDS STEC 7 (eae) Tq	AOAC 071303		Positive/Negative	
		Listeria spp.	Assurance GDS	AOAC 070701			
		E. coli O157:H7	Assurance GDS E.coli O157:H7 Tq	AOAC 2005.04			
		Listeria monocytogenes	Assurance GDS	Listeria monocytogenes Tq			AOAC 070702
		Salmonella	Assurance GDS	AOAC 2009.03			
	Food Products	Total Coliform E. coli	Total Coliform	SimPlate Color Indicator	D.L. = 1 CFU/g		
			E. coli	AOAC 2005.03			
		Total Plate Count	SimPlate Color Indicator	AOAC 2002.07			
		Staphylococcus Aureus	Petrifilm Staph Express Plate	AOAC 2003.07			
Yeast & Mold		Petrifilm Plate Count	AOAC 997.02				
E. coli O157:H7		Assurance GDS E.coli O157:H7 Tq	AOAC 2005.04	Positive/Negative			
Listeria spp.		Assurance GDS Listeria spp. Tq	AOAC 070701				
Listeria monocytogenes		Assurance GDS	Listeria monocytogenes Tq			AOAC 070702	



Certificate of Accreditation: Supplement

Altamaha Laboratories a subsidiary of Tindall Enterprises, Inc.

829 S.W. Central Avenue, Blackshear, GA 31516
Contact Name: Ana Rocks Phone: 912-449-0999

Accreditation is granted to the facility to perform the following testing

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Biological ^F	Food Products	Salmonella	Assurance GDS Salmonella Tq AOAC 2009.03	Positive/Negative
		Shiga Toxin E.coli	Assurance GDS STEC 7 (eae) Tq AOAC 071303	
Environmental ^F	Potable Water	pH	EPA 150.1	N/A
		Nitrate	SM 4500-N03 E	D.L. = 0.05 mg/L
		Nitrate as N	SM 4500-N03 D	D.L. = 0.1 mg/L
	Non-Potable Water	Biochemical Oxygen Demand	SM 5210B	D.L. = 2.0 mg/L
		Chemical Oxygen Demand	HACH 8000	D.L. = 20 mg/L - mid range D.L. = 3 mg/L - low range
		Ammonia	SM 4500-NH3 D	D.L. = 0.1 mg/L
		Nitrate as N	SM 4500-N03 D	
		TKN	SM 4500-NORGC+NH3 D	
		Orthophosphate as P	HACH 8048	
		Phosphorous, Total	HACH 8190	D.L. = 0.05 mg/L
		Nitrate	SM 4500-N03 E	
		Residue-nonfilterable	SM 2540 D	
		Oil & Grease-HEM	EPA 1664	D.L. = 5 mg/L
		Total Petroleum Hydrocarbons (SGT-HEM)	EPA 1664	
		Enterococci	IDEXX Enterolert/SM 9230 B	D.L. = 1 MPN/100 mL
		Total Coliform	SM 9223 B	
		Fecal Coliform E. coli	18-hr Enzyme Substrate	
		pH	EPA 150.1	N/A
		Fecal Coliform	SM 9221 B/E	D.L. = 1.8 mg/L
		E. coli	SM 9221 F	

- The presence of a superscript F means that the laboratory performs testing of the indicated parameter at its fixed location. Example: Outside Micrometer^F would mean that the laboratory performs this testing at its fixed location.