

# East Lake Tarpon Special Fire Control District



## *SOP 518 Suction Unit*

**Implementation  
Date: 11/2000**

**Revision Date(s): 09/2004**

**Reviewed Date(s):**

**Forms or Attachments: None**

### Daily Check

1. Check for all accessories
2. Test suction unit. (Rotate LP12 batteries as necessary.)
3. Be sure suction unit is clean and ready for use.

### After Each Use

1. Disconnect all tubing.
2. Dispose of contaminated canister/hoses.
3. Assemble with new canister.