

MRI info – Siemens Avanto 1.5T TIM

The compressor is out, the magnet is warm and there is no helium.

1. YOM: 2004 (?)
2. Magnet Type – Avanto 1.5
3. Computer type: HPZ400 model #10660928
4. Gradient: K2259/1500V/550A Gradient Coil A505 Model #07563955 K2259
5. Helium Level: 0.0%
6. Bore size: 60cm
7. # of Channels: 18 channels
8. Cold head replaced: unknown
9. Last PM: unknown
10. List of Coils:
 - a. Body Matrix A Trim Coil model #07579555
 - b. Head Coil model #07577732
 - c. Siemens 8 channel foot & ankle array assembly #108542
 - d. Neck matrix A Trim coil model #07577906
 - e. Shoulder array coil set assembly #104157 part #7580025
 - i. Shoulder array coil assembly #100214 part #44548
 - ii. Shoulder array coil assembly #100213 part#7580025
 - f. CP small flex coil model #05512038
 - g. CP large flex coil model #05512053
 - h. CP extremity coil model #07579472
 - i. Spine matrix A Trim coil model #07579340
11. Is it still under service contract? No
12. Location within facility: Southeast corner of building 1st floor
13. Loading dock: No
14. Facility will be responsible to make clear pathways for removal.







