

## PERRY JOHNSON LABORATORY ACCREDITATION, INC.

# Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

#### Altamaha Laboratories a subsidiary of Tindall Enterprises, Inc. 829 S.W. Central Avenue, Blackshear, GA 31516

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

### ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

#### **Biological and Chemical Testing** (As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen President

Perry Johnson Laboratory Accreditation, Inc. (PJLA) 755 W. Big Beaver, Suite 1325 Troy, Michigan 48084 Initial Accreditation Date: Issue Date: Expiration Date: February 27, 2013 February 28, 2023 March 31, 2025 Accreditation No.: Certificate No: 73811 L23-167

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: <u>www.pjlabs.com</u>



Certificate of Accreditation: Supplement

#### Altamaha Laboratories a subsidiary of Tindall Enterprises, Inc.

829 S.W. Central Avenue, Blackshear, GA 31516 Contact Name: Ana Rocks Phone: 912-449-0999

Accreditation is granted to the facility to perform the following testing:				
FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Biological <sup>F</sup>	Potable Water	Total Coliform	SM 9221 D	Presence/Absence
		E. coli	SM 9221 F	
		Total Coliform	SM 9223 B	
		E.coli	(Modified Colitag)	
	Environmental	Total Coliform	SM 9223 B	D.L. = 1 MPN/100 mL
	Non-Potable Water	Fecal Coliform E. coli	Colilert-18	
		Enterococci	IDEXX Enterolert®	D.L. = 1 MPN/100 mL
		Total Coliform	SM 9223 B	
		Fecal Coliform <i>E. coli</i>	Colilert-18®	
		Fecal Coliform	SM 9221 B/E	D.L. = 1.8 mg/L
		E. coli	SM 9221 F	
Chemical <sup>F</sup>	Potable Water	Nitrate + Nitrite as N	SM 4500-N03 E	D.L. = 0.05 mg/L
		Nitrate as N	SM 4500-N03 D	D.L. = 0.1 mg/L
	Non-Potable Water	Biochemical Oxygen Demand	SM 5210 B	D.L. = 2.0 mg/L
		Chemical Oxygen Demand	HACH 8000	D.L. = $20 \text{ mg/L} - \text{mid range}$ D.L. = $3.0 \text{ mg/L} - \text{low range}$
		Nitrate as N	SM 4500-NO3 D	D.L. = 0.1 mg/L
		Nitrate + Nitrite as N	SM 4500-NO3 E	D.L. = 0.05 mg/L
		Ammonia	SM 4500-NH3 D	D.L. = 0.1 mg/L
			EPA 350.1	D.L. = 0.02 mg/L
		TKN	SM 4500-Norg C + SM 4500 NH3 D	D.L. = 0.1 mg/L
			EPA 351.2	D.L. = 0.5 mg/L
		Orthophosphate as P	HACH 8048	D.L. = 0.1 mg/L
			EPA 365.1	D.L. = 0.05 mg/L
		Phosphorous, Total	HACH 8190	D.L. = 0.1 mg/L
			EPA 365.1	D.L. = 0.05 mg/L
		Residue-nonfilterable	SM 2540 D	D.L. = 5.0 mg/L
		Oil & Grease-HEM	EPA 1664	
		Total Petroleum	EPA 1664	
		Hydrocarbons (SGT-HEM)		

ccreditation is granted to the facility to perform the following testing:

1. The presence of a superscript F means that the laboratory performs testing of the indicated parameter at its fixed location. Example: Outside Micrometer<sup>F</sup> would mean that the laboratory performs this testing at its fixed location.