



Nationwide Gage Calibration Inc.

Certificate of Calibration

ISO Certificate



Sample Company
1234 Happy Lane
Chicago IL 60651

Date Calibrated: 8/17/2011
Date Due: 8/17/2012

Purchase Order # 76259

Test # 11-45956

Instrument: Mitutoyo Digital Caliper
Serial # 45789
Accuracy: Manufacturer's Specifications

Model # 500-196
Gage ID # Cal 0124

Calibration Results

Function Tested: Inside & Outside Diameter
Depth
Step

As Found Conditions: In Tolerance

As Left Conditions: In Tolerance

Calibration Procedure(s) CAL-11-ISOCAL-010

Calibration Environmental Conditions: 68 degrees F. ($\pm 3^\circ$) 45% R.H. ($\pm 10\%$)

Technical Remarks: Cleaned & Lubricated

We hereby certify that this instrument has been tested with standards & equipment calibrated with traceability to the National Institute of Standards & Technology (N.I.S.T.). The determination of accuracy (observational & mechanical error) is dependent upon the condition of the instrument being calibrated. The results are only applicable for the date of calibration noted. This instrument was calibrated in accordance with customer supplied requirements, NGC's calibration procedure CAL-11-ISOCAL-010, and manufacturer's specifications maybe used as reference if applicable. All calibrated instruments are processed under the NGC Quality Management System. This certificate shall not be reproduced without the written consent of Nationwide Gage Calibration Inc.

Calibration Standards

<u>Gage ID #</u>	<u>Standard</u>	<u>Traceability #</u>	<u>Date Calibrated</u>	<u>Date Due</u>
B12	Gage Block Set	11-B12	5/23/2011	5/23/2012
E9	Caliper Master	10-E9	9/20/2010	9/20/2011
K12	Surface Plate	08-K12	11/11/2008	9/11/2011

Technician: _____
Ronald Stade

Approved By: _____