

Date	Date	Date	Date	Date	Date	Date
Aliquot 1	Aliquot 2	Aliquot 3	Aliquot 4	Aliquot 5	Aliquot 6	Aliquot 7
0.048	0.049	0.051	0.051	0.052	0.05	0.05

Std Dev
0.001345

MDL
0.004224

Standard Deviation
Number 1 = Aliquots (7)

MDL
Standard Deviation x 3.14

Copper RL = 0.050 mg/L
