



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017
& ANSI/NCSL Z540-1-1994

QUALITY TECH SERVICES
132 Galan Drive
Byron, GA 31008
Michael Schraufnagel Phone: 478 654 6690

CALIBRATION

Valid To: April 30, 2026

Certificate Number: 2921.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1, 6}:

I. Dimensional

Parameter/Equipment	Range	CMC ^{2, 4, 5} (\pm)	Comments
Machine Tools, CMMs, CNCs, Measuring Machines, Scales and Encoders ³ –			
Linear Displacement Accuracy	Up to 3200 in	2 μ in/in (2 μ m/m)	Laser systems, XL, XM
Linear Diagonals	Up to 3200 in	2 μ in/in (2 μ m/m)	
Straightness and Squareness – Measure ³	Up to 160 in (Using Short Range Optics – XL only)	$[150 + (0.005S + 20 + 0.5F^2)] \mu$ in	Laser systems, XL, XM
Squareness XL Only	Up to 1200 in (Using Long Range Optics – XL only)	$[140 + (0.025S + 200 + 0.05F^2)] \mu$ in	

Parameter/Equipment	Range	CMC ^{2, 4, 5} (\pm)	Comments
Angular Displacement – CNC, CMM, Inclinometers and Levels ³	(-10 to 10) $^{\circ}$	[0.38 + (0.006 <i>A</i> + 0.1 + 0.007 <i>F</i>)] arc seconds [(0.02 % <i>A</i>) + 0.5 + 0.1 <i>M</i>] μ m/m	Laser systems, XL, XM w/ special optics with measured angular factor (XL only)
Angle – Measure ³	Up to 40 000 μ m/m	3.4 μ m/m (0.7 arc seconds)	Wyler electronic levels, wireless
Rotational Positioning – Rotary Encoders, Inclinometers/Levels, Rotary Indexes/Tables ³	(-360 to 360) $^{\circ}$	[2.2 + (0.006 <i>A</i> + 0.1)] arc seconds [(0.02 % <i>A</i>) + 0.5 + 0.1 <i>M</i>] μ m/m	RX10, Ultradex rotary calibrator, wireless XR20-W w/ special optics with measured angular factor (XL only)
Circular Interpolation Volume – CMMs, CNC and Machine Tools ³	(50 to 600) mm	[2.4 + (0.4 % <i>V</i>)] μ m	QC20-W ball bar, wireless

¹ This laboratory offers commercial calibration and field calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g., resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ In the statement of CMC, *S* is displayed straightness reading in μ m; *F* is distance traveled in feet; *A* is the displayed angle in arc seconds; *A* is the angle measured in μ radians; *M* is the linear measurement between angular interferometer and reflection in meters; and *V* is the radial circularity variation reading in μ m.

⁵ The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.

⁶ This scope meets A2LA's *P112 Flexible Scope Policy*.





Accredited Laboratory

A2LA has accredited

QUALITY TECH SERVICES

Byron, GA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NC SL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 4th day of March 2024.

A blue ink signature of Mr. Trace McInturff, written over a horizontal line.

Mr. Trace McInturff, Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 2921.01
Valid to April 30, 2026

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.