



Maine Center for Disease
Control and Prevention
An Office of the
Department of Health and Human Services

Paul R. LePage, Governor

Mary C. Mayhew, Commissioner

Department of Health and Human Services
Health and Environmental Testing Laboratory
211 State Street
12 State House Station
Augusta, Maine 04333-0012
Tel: (207) 287-2727; Fax: (207) 287-6832
TTY: 1-800-606-0215

TOWN OF BALDWIN
534 PEQUAWKET TRAIL
WEST BALDWIN ME 04091

Fax#:

Logged: 5/24/2016 8:27:00AM
Folder/ Invoice # L015813

Office Use Only:
Do Not Bill
WALK-IN

Project Name: SAND POND

No. of Samples in Folder: 1

L015813001

Released: 5/26/2016
Case #:

CERTIFICATION

The HETL hereby certifies that all test results for this report were analyzed by the method listed and meet all NELAC requirements, unless otherwise noted.

Kenneth G. Pote, PhD., Director

Richard French, Quality Assurance Officer

If we can be of further assistance to you, Please Call us at 287-1716

Approved by:

Christopher Montagna
Inorganics Supervisor/Chemist III

Continued from Previous Page

HETL Folder Number: L015813

HETL Sample Number: L015813001

Description: BEACH NP-H20

Matrix: NP-H20

Sample Point:

Sampler: OLIN THOMAS

Sample Date: 5/23/2016

Time: 13:00:00

Analyte	Result	Units	Qualifier	RL	Method	Analyst	Analysis Datetime
E. coli	<1	MPN/100 ml	J	1	9223 B	johnc	05/24/2016, 11:15

Comment

Attached By JCC

Date 5/25/2016 12:00:00AM Time 17:26

Comment : E.coli result is approximate because sample exceeded 8-hour holding time for non-potable water.

Continued from Previous Page

HETL Folder Number: L015813

Units & Measurement

"mg/L" = Milligrams per liter;

"ug/L" = Micrograms per Liter;

"mg/Kg" = Milligrams per Kilogram;

"ug/Kg" = Micrograms per Kilogram;

"PPM" = Parts per Million;

"NTU" = Nephelometric Turbidity Units;

All solid results on a "Dry Weight" basis

NC = Not confirmed NQ = Not Quantitated NA = Not Analyzed J = Approximately U = Undetected R = Rejected

RL-Reporting Limit, the lowest concentration which can be reliably reported on a routine basis

"<" = Less than ">" = Greater than

Note: Samples tested for analytes requiring cooling as preservation received above the preservation temperature may affect the reported result.

Results below the advisory limit, including < and K are considered satisfactory for that parameter.

Disclaimer

Your report consists of the number of pages listed on the cover page. Any attachments after the last numbered page are for informational purposes only and not part of the formal report.

The results in this report are for the submitted sample(s) only.

This report shall not be reproduced, except in full, without written permission from the Maine Health and Environmental Testing Laboratory.

